



® Management Procedure 2574
Revision: B
Date Issued: March 15, 2019
Date Revised: September 16, 2019

Calibration Procedure

DeFelsko Corporation

PosiTest HHD

High voltage Holiday Detector

Table of Contents

1	Introduction and UUC Performance Requirements.....	2
	Table 1-1 Measurement Ranges	2
2	Measurement Standards and Support Equipment Performance Requirements.....	2
	Table 2-1 UUC Accuracy Requirements and Description.....	2
	Table 2-2 Minimum use specification	2
	Table 2-3 Actual Equipment Specification.....	2
	Table 2-4 Calibration Environmental and Warm-Up Requirements	3
3	Uncertainty Calculation	3
4	Preliminary Operations.....	4
5	Calibration Process	4
	Table 5-1.....	4
6	Performance Requirements.....	5
	Table 6-1 Performance Requirements and Calibration Data for PosiTest HHD	5

1 Introduction and UUC Performance Requirements

1.1 This procedure describes the calibration of DeFelsko Corporation PosiTest HHD with the following specification:

Table 1-1 Measurement Ranges

Probe	Measurement Range
500 – 35,000V	±5%

1.2 The unit being calibrated will be referred to as the UUC (Unit-Under-Calibration).

2 Measurement Standards and Support Equipment Performance Requirements

2.1 The UUC accuracy requirements are based upon the published UUC performance specifications.

2.2 The test uncertainty ratio (TUR) applied in this Calibration Procedure is 4:1 unless otherwise stated.

2.3 The Minimum-Use-Specifications are the minimum test equipment specifications required to meet all the UUC accuracy requirements and the test uncertainty ratio applied.

Table 2-1 UUC Accuracy Requirements and Description

UUC	Performance Specifications	Test Method
HHD	500 – 35,000 V ± 5% of reading	Digitizer and High Voltage Probe

Table 2-2 Minimum use specification

Range	Accuracy
500 – 35,000V	± 1.25%

Table 2-3 Actual Equipment Specification

Equipment Generic Name	Range	Accuracy	Manufacturer / Model #'s Applicable
Digitizer	0 to 2V [†] >2 to 32V [†] >32 to 64V [†] >64 to 128V [†]	± (0.10% of reading + 0.12% of range) ± (0.10% of reading + 0.12% of range) ± (0.10% of reading + 0.30% of range) ± (0.10% of reading + 0.20% of range)	Keysight, L4532A
High Voltage Probe	N/A	1000:1 attenuation ± 0.5% ^{††}	Tektronix, P6015A

[†]With the use of the high voltage probe the effective digitizer range is multiplied by 1000.

†† DeFelsko restricts the manufacturer’s specification from 3% to 0.5% by requiring tighter calibration limits.

Caution: The instructions in this Calibration Procedure relate specifically to the equipment and conditions listed in Section 2. If other equipment is substituted, the information and instructions must be interpreted accordingly.

Table 2-4 Calibration Environmental and Warm-Up Requirements

Measurement Standards & Support Equipment Environmental Requirements:	Temperature: $23 \pm 5^{\circ} \text{C}$. Relative Humidity: Less than 95%
Measurement Standards & Support Equipment Warm-up and Stabilization Requirements:	60 minutes

3 Uncertainty Calculation

3.1 The high voltage probe is used to reduce the voltage output from the PosiTest HHD to levels that are within the Digitizer’s input range. When calculating the error associated with the digitizer, the input value is the HHD output divided by 1,000. The calculated digitizer error then needs to be multiplied by 1,000 before being combined with the voltage probe error.

3.2 At a PosiTest HHD output of 500V, digitizer input of 2V :
 High Voltage Probe accuracy @ 500V = $500 * 0.5\% = 2.5\text{V}$
 Digitizer accuracy @ 2V range = 0.10% of reading + 0.12% of range
 $= (0.5 * 0.10\%) + (2 * 0.12\%)$
 $= 0.0029\text{V}$
 The combined accuracy using a sum of squares = $(\text{probe}^2 + \text{digitizer}^2)^{0.5}$
 $= (2.5^2 + (0.0029 * 1000)^2)^{0.5}$
 $= (6.25 + 8.41)^{0.5}$
 $= 4\text{V}$
 $\text{TUR} = (500 * 5\%) / 4\text{V} = 6$

3.3 At a PosiTest HHD output of 5,000V, digitizer input of 32V :
 Probe accuracy @ 5,000V = $5000 * 0.5\% = 25\text{V}$
 Digitizer accuracy @ 32V range = 0.10% of reading + 0.12% of range
 $= (5 * 0.10\%) + (32 * 0.12\%)$
 $= 0.0434\text{V}$
 The combined accuracy using a sum of squares = $(\text{probe}^2 + \text{digitizer}^2)^{0.5}$
 $= (25^2 + (0.0434 * 1000)^2)^{0.5}$
 $= (625 + 1884)^{0.5}$
 $= 50\text{V}$
 $\text{TUR} = (5,000 * 5\%) / 50\text{V} = 5$

3.4 At a PosiTest HHD output of 20,000V, digitizer input of 64V :
 Probe accuracy @ 20,000V = $20000 * 0.5\% = 100\text{V}$
 Digitizer accuracy @ 64V range = 0.10% of reading + 0.30% of range

$$= (20*0.10\%) + (64*0.30\%)$$

$$= 0.212V$$

The combined accuracy using a sum of squares = $(\text{probe}^2 + \text{digitizer}^2)^{0.5}$

$$= (100^2 + (0.212*1000)^2)^{0.5}$$

$$= (10000 + 44944)^{0.5}$$

$$= 234V$$

$$\text{TUR} = (20,000*5\%) / 234V = 4$$

3.5 At a PosiTest HHD output of 35,000V, digitizer input of 128V :

Probe accuracy @ 35,000V = $35000 * 0.5\% = 175V$

Digitizer accuracy @ 128V range = 0.10% of reading + 0.20% of range

$$= (35*0.10\%) + (128*0.20\%)$$

$$= 0.291V$$

The combined accuracy using a sum of squares = $(\text{probe}^2 + \text{digitizer}^2)^{0.5}$

$$= (175^2 + (0.291*1000)^2)^{0.5}$$

$$= (30625 + 84681)^{0.5}$$

$$= 340V$$

$$\text{TUR} = (35,000*5\%) / 340V = 5$$

4 Preliminary Operations

Note: Review the entire document before starting the calibration process.

4.1 Connect the High Voltage Probe to the UUC output and connect the digitizer to the High Voltage Probe output.

4.2 Connect the UUC to ground and turn on the unit.

5 Calibration Process

Note: Whenever the test requirement is not met, verify the results of each test and take corrective action before proceeding.

5.1 Adjust the UUC output to 500V and the Digitizer settings per table 5-1.

Table 5-1

UUC set point (V)	Digitizer Range Setting (V)	Scale (V)	Trigger* (V)
500	2	0.25	0.1
5,000	32	2	1
20,000	64	8	1
35,000	128	16	1

* Trigger must be set to Positive Edge Mode.

- 5.2 Activate the UUC Detection Mode by pressing and holding the power button and then pulling and holding the trigger. Once the Detection Mode is started the power button can be released but continue holding the trigger for at least 5 seconds. Release the trigger to exit Detection Mode. Note the Vmax measurement from the digitizer and the voltage displayed on the UUC.
- 5.3 Repeat step 5.2 for a total of ten readings. Record the average of the ten readings in Table 6-1 then repeat steps 5.1 & 5.2 for all of the set points in Table 5-1.

6 Performance Requirements

Note: The technician will collect the data needed to complete the appropriate table below. Do not write in this procedure.

Table 6-1 Performance Requirements and Calibration Data for PosiTest HHD

UUC Set Point (V)	UUC Reading (V)	Digitizer Reading (V)	Min Reading Allowed (V) ❶	Max Reading Allowed (V) ❷
500				
5,000				
20,000				
35,000				

❶ Calculation: (Digitizer reading times 1000) times 0.95. Round up to the nearest 1V.

❷ Calculation: (Digitizer reading times 1000) times 1.05. Round down to the nearest 1V.

