PosiTest[®] HHD Verifier

Standalone volt meter (crest meter, jeep meter) for verifying the accuracy and operation of ALL <u>pulsed</u> type DC holiday detectors

Features...

- Ideal for field/laboratory verification of pulsed DC holiday detectors
- Wide measurement range: 0.5 35 kV
- Simple two step operation
- Lightweight/Portable requires no external power source
- Certificates of Calibration showing traceability to NIST included
- Two year warranty
- Made in USA

EASY OPERATION

- Step 1. Ensure that the ground lead is plugged into the PosiTest HHD Verifier and that the other end of the ground lead is connected to the ground lead of the high voltage holiday detector that is being verified.
- Step 2. Turn the **PosiTest** *HHD Verifier* on by pressing and releasing the power button. Once powered on, the verifier continuously monitors the voltage on the probe end of the verifier. The verifier will update the display when the probe is touching an exposed electrode of a holiday detector. The voltage displayed on the verifier should match the voltage displayed by the detector within the combined tolerance of both units.

DEVICE COMES COMPLETE with verifier, 1.8 m (6') ground wire, 3 AAA batteries, instructions, Long Form Certificate of Calibration traceable to NIST, two year warranty.

SIZE: 43 x 5 x 5 cm (17" x 2" x 2")

WEIGHT: 476 g (1.05 lbs.)

C	vormer Vormer 3.7 J. pbc.w			
		9	_	
	~			
			0	

Specifications			
Voltage Range	0.5 – 35 kV		
Resolution	0.5 kV – 1 kV: 10 Volt Steps 1 kV – 35 kV: 100 Volt Steps		
Accuracy	<u>+</u> 5%		
Operating Temperature	-20° to 60°C (-4° to 140°F)		

Conforms to ASTM D4787/D5162/G62, ISO 29601, NACE SP0188/ SP0274/SP0490, AS3894.1



DeFelsko Corporation • 800 Proctor Ave., Ogdensburg, New York 13669-2205 USA **Toll Free 1-800-448-3835** • Tel: +1-315-393-4450 • Fax: +1-315-393-8471 www.defelsko.com • E-mail: techsale@defelsko.com



© 2023 DeFelsko Corporation USA. All Rights Reserved. Technical Data subject to change without notice. Printed in USA. HHDV.v.LW-W2306