

DeFelsko Corporation 800 Proctor Avenue Ogdensburg, New York 13669-2205 USA

Batch Test Certificate

Product: Bresle Method Latex Patch

Batch: ABC123
Best Before: Oct 2018
Test Date: May 16, 2017

DeFelsko Corporation certifies that the identified product has been tested in accordance with the following standards.

Standard Reference	Description	Criteria	Result
ISO 8502-6	Patch Thickness	$1.5 \text{mm} \pm 0.3 \text{mm}$	1.5 mm
Section 4.1			
ISO 8502-6	Adhesive Rim	> 5mm	6.0 mm
Section 4.1	Width		
ISO 8502-6	A-1250	$1250 \text{mm}^2 + 13 \text{mm}^2$	1250 mm^2
Section 4.1 Table 1	Compartment Area		
ISO 8502-6	Leak Test	8 of 12 contain 14.9ml	12 Pass
Annex A		\pm 0.1ml for 20 minutes	
ISO 8502-9	Ionic	$<7 \text{mg/m}^2$	$< 5 \text{ mg/m}^2*$
Section 5.3 Note	Contamination	_	

*Uncertainty $\pm 1.19 \text{ mg/m}^2$

Test Performed by:	Sample		
		Technician	

Avoid storage in sunlight or extended periods above 30°C (86°F).

This certificate shall not be reproduced, except in full, without the written approval of DeFelsko Corporation.