

Batch Test Certificate

Product: Bresle Method PosiPatch
Code: SSTPATCHP5
Batch: ABC123
Test Date: May 16, 2017

DeFelsko Corporation certifies that the identified product has been tested in accordance with the following standards.

Standard Reference	Description	Criteria	Result
ISO 8502-6 Section 4.1 Table 1	A-1250 Compartment Area	$1250\text{mm}^2 \pm 13\text{mm}^2$	1250 mm^2
ISO 8502-9 Section 5.3 Note	Ionic Contamination	$< 7\text{mg}/\text{m}^2$	$< 3\text{ mg}/\text{m}^2*$

*Uncertainty $\pm 1.19\text{ mg}/\text{m}^2$

Test Performed by: Sample
Technician

This certificate shall not be reproduced, except in full, without the written approval of DeFelsko Corporation.